



Certificate of Compliance

Certificate: 1557657 (LR 44340C)

Master Contract: 162029

Project: 2524963

Date Issued: October 22, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: Thomas Smith

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Handheld Digital Multimeters, Models Fluke 83 Series V, 87 Series V, 87V Ex, and 88 Series V; operated by a 9 V battery.

Note: The above models are Pollution Degree 2, Measurement Category 1000 V Cat III; 600 V Cat IV.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Std. No. 61010-1 (3rd ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st ed.) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits



Certificate: 1557657 (LR 44340C)

Master Contract: 162029

Project: 2524963

Date Issued: October 22, 2012

Reference Standards

IEC 61010-1:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-030:2010 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits