

Manual Supplement

Manual Title:	71X Series Calibration	Supplement Issue:	2
Part Number:	686540	Issue Date:	10/12
Print Date:	January 1998	Page Count:	1
Revision/Date:	11, 3/12		

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On page 49, under *mA Loop Power Verification*, replace step 2 with:

2. Apply 1.2 k Ω to the mA jacks from the decade box or other resistor source. The display must read >19 mA but <OL.

On page 54, under *mA Loop Power Verification (718 Only)*, replace step 2 with:

2. Apply 1.2 k Ω to the mA jacks from the decade box or other resistor source. The display must read >19 mA but <OL.

Change #2

On page 86, Table 71,

Change:

MP6	30G, 100G, 300G Pump Assembly	2558508	718	1
-----	-------------------------------	---------	-----	---

To:

MP6	Fluke-718, 30G/100G, Pump with clean out	2558508	718	1
MP6	Fluke-718, 300G, Pump with clean out	2632057	718	1